



Ref. Certif. No.

DK-127541-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

SWITCHING POWER SUPPLY

Name and address of the applicant

DONGGUAN AOYUAN ELECTRONICS TECHNOLOGY CO LTD  
FACTORY A7 ZHEN'AN SCI-TECH INDUSTRIAL PARK  
CHANG'AN TOWN DONGGUAN GUANGDONG 523872  
CHINA

Name and address of the manufacturer

DONGGUAN AOYUAN ELECTRONICS TECHNOLOGY CO LTD  
FACTORY A7 ZHEN'AN SCI-TECH INDUSTRIAL PARK  
CHANG'AN TOWN DONGGUAN GUANGDONG 523872  
CHINA

Name and address of the factory

DONGGUAN AOYUAN ELECTRONICS TECHNOLOGY CO LTD  
FACTORY A7 ZHEN'AN SCI-TECH INDUSTRIAL PARK  
CHANG'AN TOWN DONGGUAN GUANGDONG 523872  
CHINA

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input: 100-240V~, 50/60Hz, 3.5A Max.  
Output: See test report for details.

Trademark (if any)

None

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

AYS218Y-xxx  
 Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

**Additionally evaluated to:** EN IEC 62368-1:2020, EN IEC 62368-1:2020/A11:2020  
National Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 62368-1:2018

As shown in the Test Report Ref. No. which forms part of this Certificate

22SBCS02003 01111 issued on 2022-05-06

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/nbcnames](http://www.ul.com/nbcnames)

Date: 2022-05-13

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-127541-UL**

**Additional Model Detail(s):**

AYS218Y-xxx, xxx = 038-042, 3 digits, indicates 10 times of the output voltage in V, minimum rising step is 0.1V. e.g.:038=3.8V, 042=4.2V.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2022-05-13

Signature:

Jan-Erik Storgaard